

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4577

Name: RITTER 1  
Sterilizer ID: V2178198  
05/23/2024 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.1 F 0.0  
2:00 142.8 F 1.9  
4:00 157.3 F 4.8  
6:00 182.5 F 7.9  
8:00 199.6 F 12.3  
10:00 209.5 F 3.9  
12:00 221.7 F 7.8  
14:00 231.7 F 11.9  
16:00 231.7 F 7.9  
18:00 243.2 F 13.1  
20:00 242.0 F 11.4  
22:00 253.7 F 17.3  
24:00 264.2 F 23.5  
26:00 272.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.2  
2:00 272.5 F 28.9  
3:00 271.7 F 28.4  
4:00 271.7 F 28.7

Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 200.7 F 0.0  
10:00 209.9 F 0.0  
15:00 211.7 F 0.0  
20:00 214.3 F 0.0  
25:00 216.6 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:12  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:04:52  
CYCLE COMPLETE

05/23/2024 03:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_