
Midmark M11 - v1.0.0
Total Cycles: 2573

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.7 F 0.0
2:00 80.9 F 0.0
4:00 87.8 F 0.0
6:00 114.7 F 0.1
8:00 149.5 F 2.6
10:00 175.0 F 6.1
12:00 193.9 F 9.9
14:00 207.6 F 8.7
16:00 219.6 F 6.4
18:00 230.5 F 10.3
20:00 234.9 F 8.5
22:00 241.7 F 11.8
24:00 239.0 F 9.9
26:00 252.0 F 16.2
28:00 263.0 F 22.7
30:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 271.9 F 28.9
2:00 272.0 F 28.5
3:00 271.6 F 28.7
4:00 272.2 F 28.8

Min 271.2 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 209.3 F 0.0
10:00 197.6 F 0.0
15:00 198.4 F 0.0
20:00 197.5 F 0.0
25:00 205.0 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:08
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

11/17/2022 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____