
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008323
9:54 19/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:35
STERILIZING 3:37
136.0 C 220kPa 3:37
136.4 C 223kPa 4:07
136.3 C 222kPa 4:37
136.4 C 223kPa 5:07
136.1 C 219kPa 5:37
136.3 C 220kPa 6:07
136.4 C 223kPa 6:37
136.0 C 217kPa 7:07
136.0 C 217kPa 7:07
VENTING 7:07
AIR DRYING 8:00
CYCLE COMPLETE 22:55

Digital Signature #
C972B01CD60A2EC3
