

-----  
Midmark M11 - v1.0.5  
Total Cycles: 179

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2021 06:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 74.2 F 0.0  
4:00 83.5 F 0.1  
6:00 122.4 F 1.1  
8:00 152.9 F 4.0  
10:00 178.3 F 7.4  
12:00 193.7 F 11.2  
14:00 208.9 F 3.4  
16:00 223.7 F 7.3  
18:00 232.5 F 11.4  
20:00 230.7 F 7.5  
22:00 243.1 F 12.8  
24:00 241.2 F 11.3  
26:00 253.2 F 17.4  
28:00 263.8 F 23.7  
30:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.7 F 29.3  
2:00 272.3 F 29.5  
3:00 272.1 F 29.0  
4:00 272.7 F 29.7

Min 271.6 F 28.7  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 199.5 F 0.0  
10:00 199.9 F 0.0  
15:00 198.6 F 0.0  
20:00 193.8 F 0.0  
25:00 189.7 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 30:07  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:08:16  
CYCLE COMPLETE

07/08/2021 07:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_