
Midmark M11 - v1.0.5
Total Cycles: 4141

Name: RITTER2
Sterilizer ID: V2313228
11/20/2024 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 153.7 F 1.6
4:00 174.3 F 5.3
6:00 203.0 F 10.0
8:00 222.5 F 6.9
10:00 226.1 F 9.9
12:00 232.6 F 8.6
14:00 243.8 F 12.9
16:00 242.4 F 11.9
18:00 255.2 F 18.6
20:00 266.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 273.1 F 29.8
2:00 272.3 F 30.0
3:00 271.6 F 29.3
4:00 273.0 F 29.8

Min 271.0 F 28.8
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 218.6 F 0.0
10:00 194.4 F 0.0
15:00 188.2 F 0.0
20:00 177.3 F 0.0
25:00 175.9 F 0.0
30:00 174.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

11/20/2024 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____