
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000212
8:40 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:25
STERILIZING 4:59
136.0 C 228kPa 4:59
136.4 C 232kPa 5:29
136.4 C 232kPa 5:59
136.3 C 233kPa 6:29
136.4 C 233kPa 6:59
136.3 C 232kPa 7:29
136.4 C 233kPa 7:59
136.1 C 230kPa 8:29
136.1 C 230kPa 8:29
VENTING 8:30
AIR DRYING 9:34
CYCLE COMPLETE 24:18

Digital Signature #
C810A93DD65F15C6
