
Midmark M11 - v1.0.0
Total Cycles: 5063

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2024 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.7 F 0.0
2:00 79.7 F 0.0
4:00 97.2 F 0.1
6:00 129.0 F 0.2
8:00 152.3 F 0.6
10:00 180.3 F 2.5
12:00 192.4 F 5.5
14:00 207.2 F 8.8
16:00 219.2 F 12.4
18:00 223.7 F 6.2
20:00 234.5 F 10.6
22:00 233.2 F 7.9
24:00 243.7 F 12.7
26:00 241.0 F 10.9
28:00 252.9 F 16.8
30:00 263.4 F 22.9
32:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.1
2:00 272.0 F 28.4
3:00 271.5 F 28.6
4:00 272.6 F 29.0

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 189.3 F 0.0
10:00 184.6 F 0.0
15:00 206.1 F 0.0
20:00 210.2 F 0.0
25:00 218.1 F 0.0
30:00 220.3 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:10:46
CYCLE COMPLETE

11/06/2024 02:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____