
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001440
9:59 07/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:28
136.0 C 220kPa 3:28
136.4 C 225kPa 3:58
136.4 C 226kPa 4:28
136.3 C 224kPa 4:58
136.3 C 224kPa 5:28
136.2 C 223kPa 5:58
136.4 C 224kPa 6:28
136.1 C 222kPa 6:58
136.1 C 222kPa 6:58
VENTING 6:58
AIR DRYING 7:48
CYCLE COMPLETE 8:19

Digital Signature #
2BBDDDB9E3A8D0BEE
