
Midmark M11 - v1.0.5
Total Cycles: 43

Name: RITTER2
Sterilizer ID: V2313228
05/26/2021 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.5 F 0.0
2:00 115.9 F 0.8
4:00 119.0 F 2.7
6:00 143.3 F 5.7
8:00 172.4 F 9.3
10:00 190.4 F 11.5
12:00 209.5 F 6.7
14:00 222.7 F 11.4
16:00 229.3 F 8.4
18:00 239.7 F 11.9
20:00 243.3 F 13.1
22:00 255.3 F 19.3
24:00 266.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.1 F 29.9
2:00 272.3 F 30.3
3:00 272.1 F 30.5
4:00 271.6 F 30.0

Min 271.5 F 29.8
Max 272.6 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 216.5 F 0.0
10:00 217.3 F 0.0
15:00 197.5 F 0.0
20:00 191.0 F 0.0
25:00 188.3 F 0.0
30:00 202.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:41
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:04:00
CYCLE COMPLETE

05/26/2021 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____