
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001810
16:10 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:39
STERILIZING 4:47
136.4 C 224kPa 4:47
136.3 C 225kPa 5:17
136.1 C 221kPa 5:47
136.4 C 225kPa 6:17
136.3 C 224kPa 6:47
136.2 C 221kPa 7:17
136.4 C 224kPa 7:47
136.3 C 225kPa 8:17
136.3 C 225kPa 8:17
VENTING 8:17
AIR DRYING 9:11
CYCLE COMPLETE 9:17

Digital Signature #
D88EFCED5A874224
