
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000792
9:34 30/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:54
136.4 C 233kPa 3:54
136.4 C 233kPa 4:24
136.4 C 233kPa 4:54
136.4 C 233kPa 5:24
136.3 C 232kPa 5:54
136.3 C 231kPa 6:24
136.4 C 231kPa 6:54
136.2 C 231kPa 7:24
136.2 C 231kPa 7:24
VENTING 7:24
AIR DRYING 8:15
CYCLE COMPLETE 8:30

Digital Signature #
CD35DFD437C5B842
