
Midmark M11 - v1.0.5
Total Cycles: 4070

Name: RITTER2
Sterilizer ID: V2313228
10/31/2024 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 71.1 F 0.0
4:00 79.9 F 0.0
6:00 118.1 F 0.1
8:00 153.0 F 0.2
10:00 182.9 F 0.3
12:00 201.4 F 2.7
14:00 210.7 F 6.1
16:00 221.8 F 9.7
18:00 231.2 F 11.8
20:00 234.7 F 9.6
22:00 238.9 F 10.2
24:00 244.9 F 13.1
26:00 243.8 F 12.3
28:00 256.4 F 18.9
30:00 267.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 28.9
2:00 272.2 F 29.2
3:00 271.9 F 28.6
4:00 272.8 F 29.4

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.7 F 0.0
5:00 176.8 F 0.0
10:00 210.8 F 0.0
15:00 207.1 F 0.0
20:00 199.8 F 0.0
25:00 196.0 F 0.0
30:00 200.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

10/31/2024 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____