

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4277

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2024 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 73.0 F 0.2  
4:00 103.2 F 1.0  
6:00 114.0 F 3.1  
8:00 141.6 F 5.6  
10:00 163.7 F 8.4  
12:00 182.3 F 11.6  
14:00 194.2 F 2.7  
16:00 209.2 F 5.3  
18:00 218.4 F 8.5  
20:00 227.7 F 12.1  
22:00 228.8 F 7.2  
24:00 240.4 F 12.0  
26:00 238.3 F 9.6  
28:00 250.2 F 15.4  
30:00 260.7 F 21.2  
32:00 270.3 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.8 F 29.0  
2:00 271.5 F 28.2  
3:00 272.8 F 29.1  
4:00 271.8 F 28.4  
  
Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 223.7 F 0.0  
10:00 235.1 F 0.0  
15:00 226.9 F 0.0  
20:00 227.3 F 0.0  
25:00 228.0 F 0.0  
30:00 225.2 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:54  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:11:17  
CYCLE COMPLETE

02/07/2024 10:41 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_