
Midmark M11 - v1.0.5
Total Cycles: 324

Name: RITTER2
Sterilizer ID: V2313228
08/31/2021 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 113.8 F 0.9
4:00 130.7 F 3.4
6:00 162.2 F 7.4
8:00 190.2 F 12.4
10:00 208.3 F 5.9
12:00 223.4 F 11.0
14:00 232.2 F 8.1
16:00 238.6 F 11.2
18:00 244.2 F 13.2
20:00 256.8 F 20.0
22:00 268.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.2
1:00 273.5 F 30.8
2:00 271.8 F 30.0
3:00 272.5 F 31.0
4:00 272.8 F 30.5

Min 271.4 F 30.0
Max 274.0 F 31.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 231.1 F 0.0
10:00 228.5 F 0.0
15:00 196.0 F 0.0
20:00 181.4 F 0.0
25:00 189.6 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:59
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:01:58
CYCLE COMPLETE

08/31/2021 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____