

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003177  
9:44 17/02/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:49  
STERILIZING 3:58  
136.5 C 220kPa 3:58  
136.4 C 223kPa 4:29  
136.4 C 224kPa 4:59  
136.1 C 218kPa 5:29  
136.1 C 222kPa 5:59  
136.3 C 223kPa 6:29  
136.4 C 222kPa 6:59  
136.2 C 221kPa 7:29  
136.2 C 221kPa 7:29  
VENTING 7:29  
AIR DRYING 8:18  
CYCLE COMPLETE 8:21

-----  
Digital Signature #  
6BC4A3FB2D47E842  
-----