
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001342
17:37 10/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 3:00
STERILIZING 4:13
136.4 C 220kPa 4:13
136.4 C 225kPa 4:43
136.2 C 223kPa 5:13
136.3 C 224kPa 5:43
136.3 C 223kPa 6:13
136.3 C 224kPa 6:43
136.2 C 223kPa 7:13
136.4 C 225kPa 7:43
136.4 C 225kPa 7:43
VENTING 7:43
AIR DRYING 8:35
CYCLE COMPLETE 9:05

Digital Signature #
4AD7E064D539D4E4
