

-----  
Midmark M11 - v1.0.5  
Total Cycles: 211

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2021 06:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.3 F 0.0  
2:00 74.0 F 0.0  
4:00 85.0 F 0.0  
6:00 103.2 F 0.2  
8:00 137.2 F 2.2  
10:00 167.2 F 5.5  
12:00 188.4 F 9.1  
14:00 206.6 F 12.8  
16:00 220.3 F 6.4  
18:00 233.4 F 10.6  
20:00 232.6 F 8.2  
22:00 241.9 F 12.3  
24:00 239.3 F 10.4  
26:00 252.1 F 16.7  
28:00 263.0 F 23.1  
30:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.8  
1:00 272.0 F 29.6  
2:00 271.8 F 29.3  
3:00 272.6 F 29.5  
4:00 272.8 F 29.9

Min 270.9 F 28.7  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
5:00 234.3 F 0.0  
10:00 223.9 F 0.0  
15:00 198.7 F 0.0  
20:00 192.6 F 0.0  
25:00 195.1 F 0.0  
30:00 194.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:12  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:09  
CYCLE COMPLETE

07/15/2021 07:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02