
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000285
11:52 02/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:10
136.0 C 229kPa 4:10
136.3 C 232kPa 4:40
136.2 C 231kPa 5:10
136.4 C 233kPa 5:40
136.4 C 233kPa 6:10
136.3 C 232kPa 6:40
136.4 C 233kPa 7:10
136.4 C 232kPa 7:40
136.4 C 232kPa 7:40
VENTING 7:40
AIR DRYING 8:35
CYCLE COMPLETE 23:35

Digital Signature #
DB866F764926CD11
