
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.3uS / 0.8ppm
CYCLE NUMBER 001525
14:41 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 5:02
STERILIZING 6:15
136.0 C 220kPa 6:15
136.5 C 223kPa 6:45
136.2 C 222kPa 7:15
136.4 C 219kPa 7:45
136.5 C 224kPa 8:15
136.2 C 222kPa 8:45
136.2 C 221kPa 9:15
136.4 C 222kPa 9:45
VENTING 9:45
AIR DRYING 10:45
CYCLE COMPLETE 40:45

Digital Signature #
C80FC8DC35E0F96E
