
Midmark M11 - v1.0.0
Total Cycles: 3601

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 186.6 F 4.2
4:00 215.6 F 9.6
6:00 223.1 F 5.9
8:00 236.2 F 11.4
10:00 235.5 F 8.8
12:00 240.5 F 9.8
14:00 247.5 F 14.0
16:00 259.7 F 20.8
18:00 270.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.0
2:00 271.5 F 28.5
3:00 272.6 F 29.2
4:00 271.7 F 28.4

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 222.9 F 0.0
10:00 235.5 F 0.0
15:00 233.1 F 0.0
20:00 226.3 F 0.0
25:00 225.7 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

08/01/2023 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____