
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001461
9:55 09/30/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:04
PRESSURIZING 6:54
STERILIZING 8:32
136.0 C 220kPa 8:32
136.4 C 223kPa 9:02
136.4 C 221kPa 9:32
136.4 C 223kPa 10:02
136.2 C 220kPa 10:32
136.3 C 222kPa 11:02
136.5 C 223kPa 11:32
136.1 C 220kPa 12:02
VENTING 12:02
AIR DRYING 13:04
CYCLE COMPLETE 15:48

Digital Signature #
64EBA092F175A121
