
Midmark M11 - v1.0.0
Total Cycles: 3014

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2023 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 197.7 F 4.3
4:00 222.9 F 10.0
6:00 229.0 F 6.2
8:00 238.4 F 11.9
10:00 237.8 F 10.0
12:00 237.7 F 8.4
14:00 251.7 F 16.1
16:00 264.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.8
2:00 271.4 F 28.1
3:00 272.5 F 29.0
4:00 271.9 F 28.4

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 229.6 F 0.0
10:00 247.8 F 0.0
15:00 244.5 F 0.0
20:00 234.6 F 0.0
25:00 223.5 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 2:22
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:28
CYCLE COMPLETE

03/16/2023 01:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____