
Midmark M11 - v1.0.5
Total Cycles: 542

Name: RITTER2
Sterilizer ID: V2313228
11/11/2021 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 163.3 F 3.2
4:00 187.5 F 8.0
6:00 209.6 F 13.3
8:00 226.2 F 8.4
10:00 235.8 F 11.9
12:00 240.0 F 11.7
14:00 240.3 F 10.9
16:00 253.5 F 17.7
18:00 264.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.7 F 29.9
2:00 271.4 F 29.1
3:00 272.3 F 29.9
4:00 271.8 F 29.3

Min 271.3 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 228.4 F 0.0
10:00 219.9 F 0.0
15:00 213.1 F 0.0
20:00 202.1 F 0.0
25:00 194.9 F 0.0
30:00 199.5 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

11/11/2021 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____