
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001921
11:54 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:07
STERILIZING 3:11
136.0 C 220kPa 3:11
136.4 C 225kPa 3:41
136.4 C 226kPa 4:11
136.3 C 224kPa 4:41
136.3 C 225kPa 5:11
136.2 C 223kPa 5:41
136.4 C 226kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:34
CYCLE COMPLETE 37:34

Digital Signature #
6DC1B9D0FFEE6C1D
