

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2718

Name: RITTER 1  
Sterilizer ID: V2178198  
12/22/2022 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.0 F 0.0  
2:00 179.3 F 2.8  
4:00 199.0 F 7.3  
6:00 218.1 F 12.6  
8:00 228.5 F 8.3  
10:00 241.2 F 12.3  
12:00 240.0 F 11.4  
14:00 238.7 F 9.7  
16:00 252.6 F 16.6  
18:00 264.4 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.3 F 28.9  
2:00 271.6 F 28.5  
3:00 272.5 F 29.0  
4:00 271.6 F 28.3  
  
Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 223.6 F 0.0  
10:00 227.6 F 0.0  
15:00 231.6 F 0.0  
20:00 233.4 F 0.0  
25:00 229.7 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

12/22/2022 12:48 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_