
Midmark M11 - v1.0.0
Total Cycles: 916

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2021 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.8 F 0.0
2:00 152.8 F 1.0
4:00 173.4 F 3.4
6:00 195.0 F 6.9
8:00 209.5 F 11.2
10:00 220.8 F 4.1
12:00 226.2 F 8.8
14:00 237.4 F 13.6
16:00 237.7 F 10.5
18:00 237.7 F 8.0
20:00 248.7 F 14.9
22:00 260.4 F 21.3
24:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.3 F 29.4
4:00 272.0 F 28.8

Min 271.3 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 208.0 F 0.0
10:00 215.2 F 0.0
15:00 218.9 F 0.0
20:00 225.7 F 0.0
25:00 228.9 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

06/18/2021 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____