

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5140

Name: RITTER 1  
Sterilizer ID: V2178198  
11/28/2024 02:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.9 F 0.0  
2:00 93.2 F 0.0  
4:00 114.8 F 0.0  
6:00 136.8 F 0.2  
8:00 168.6 F 0.6  
10:00 192.3 F 2.8  
12:00 212.4 F 6.1  
14:00 225.5 F 10.2  
16:00 233.4 F 8.1  
18:00 237.3 F 10.6  
20:00 234.9 F 7.9  
22:00 242.3 F 11.1  
24:00 246.6 F 13.5  
26:00 258.6 F 20.0  
28:00 269.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.2 F 29.0  
2:00 272.5 F 29.0  
3:00 272.1 F 28.6  
4:00 271.4 F 28.3

Min 271.3 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 203.8 F 0.0  
10:00 194.2 F 0.0  
15:00 222.7 F 0.0  
20:00 236.2 F 0.0  
25:00 228.8 F 0.0  
30:00 237.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 29:12  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:07:43  
CYCLE COMPLETE

11/28/2024 04:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_