
Midmark M11 - v1.0.0
Total Cycles: 3771

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 165.7 F 1.2
4:00 180.8 F 4.5
6:00 201.9 F 8.7
8:00 218.6 F 13.3
10:00 224.9 F 8.4
12:00 234.5 F 13.2
14:00 237.1 F 10.3
16:00 236.0 F 8.3
18:00 248.1 F 14.4
20:00 259.4 F 20.6
22:00 269.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.9 F 29.3
2:00 271.9 F 28.7
3:00 271.6 F 28.9
4:00 272.8 F 29.3

Min 271.3 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 224.7 F 0.0
10:00 233.2 F 0.0
15:00 227.1 F 0.0
20:00 226.0 F 0.0
25:00 226.7 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:55
CYCLE COMPLETE

09/14/2023 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____