

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2549

Name: RITTER2  
Sterilizer ID: V2313228  
06/22/2023 07:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 77.3 F 0.0  
4:00 96.0 F 0.0  
6:00 105.7 F 0.0  
8:00 143.8 F 2.5  
10:00 172.8 F 5.8  
12:00 202.3 F 9.5  
14:00 208.8 F 13.4  
16:00 218.2 F 6.9  
18:00 230.0 F 10.9  
20:00 230.7 F 7.4  
22:00 239.4 F 11.9  
24:00 237.4 F 9.7  
26:00 249.9 F 15.8  
28:00 260.5 F 21.8  
30:00 270.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 273.0 F 30.1  
2:00 271.9 F 29.8  
3:00 272.2 F 29.5  
4:00 272.7 F 29.9  
  
Min 271.0 F 28.9  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 200.8 F 0.0  
10:00 208.8 F 0.0  
15:00 220.5 F 0.0  
20:00 217.6 F 0.0  
25:00 220.0 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:50  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

06/22/2023 08:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_