
Midmark M11 - v1.0.0
Total Cycles: 5137

Name: RITTER 1
Sterilizer ID: V2178198
11/27/2024 05:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.6 F 0.0
2:00 112.0 F 0.1
4:00 129.5 F 0.3
6:00 164.9 F 0.5
8:00 190.1 F 1.3
10:00 209.1 F 5.1
12:00 220.3 F 9.4
14:00 235.2 F 10.1
16:00 235.6 F 10.3
18:00 235.1 F 8.0
20:00 242.5 F 11.1
22:00 247.0 F 13.6
24:00 259.7 F 20.6
26:00 270.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 29.0
2:00 272.4 F 28.7
3:00 271.6 F 28.2
4:00 271.9 F 28.6

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 184.9 F 0.0
10:00 183.5 F 0.0
15:00 188.5 F 0.0
20:00 217.4 F 0.0
25:00 211.3 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 26:51
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:19
CYCLE COMPLETE

11/27/2024 06:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____