

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2509

Name: RITTER 1  
Sterilizer ID: V2178198  
11/03/2022 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.4 F 0.0  
2:00 179.0 F 3.6  
4:00 175.8 F 8.1  
6:00 200.1 F 13.5  
8:00 223.6 F 6.8  
10:00 233.6 F 11.8  
12:00 233.3 F 8.4  
14:00 242.7 F 10.8  
16:00 245.5 F 13.0  
18:00 258.0 F 19.6  
20:00 269.3 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.7 F 29.0  
2:00 271.8 F 28.3  
3:00 272.0 F 28.8  
4:00 272.7 F 29.0

Min 271.4 F 28.1  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.6 F 0.0  
5:00 249.7 F 0.0  
10:00 224.2 F 0.0  
15:00 230.4 F 0.0  
20:00 226.4 F 0.0  
25:00 230.8 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:05  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:33  
CYCLE COMPLETE

11/03/2022 11:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_