
Midmark M11 - v1.0.0
Total Cycles: 4671

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2024 09:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 69.4 F 0.0
4:00 88.5 F 0.1
6:00 121.9 F 0.2
8:00 150.6 F 0.6
10:00 176.9 F 3.2
12:00 192.2 F 6.5
14:00 205.5 F 9.9
16:00 217.0 F 13.3
18:00 222.7 F 6.5
20:00 233.2 F 10.5
22:00 234.5 F 8.6
24:00 242.1 F 11.7
26:00 237.9 F 9.5
28:00 250.9 F 15.7
30:00 261.8 F 21.9
32:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 28.5
2:00 272.4 F 29.0
3:00 271.6 F 28.2
4:00 272.5 F 28.8

Min 271.5 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 230.8 F 0.0
10:00 224.8 F 0.0
15:00 180.3 F 0.0
20:00 179.8 F 0.0
25:00 186.2 F 0.0
30:00 188.2 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 32:33
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:47
CYCLE COMPLETE

06/24/2024 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____