
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000037
12:06 19/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:34
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:34

Digital Signature #
51A61E9F519EF1D1
