
Midmark M11 - v1.0.0
Total Cycles: 1241

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.2 F 0.0
2:00 108.5 F 0.0
4:00 132.0 F 0.1
6:00 158.0 F 1.7
8:00 185.8 F 4.7
10:00 204.0 F 7.9
12:00 217.3 F 12.1
14:00 227.4 F 6.5
16:00 239.9 F 11.7
18:00 236.6 F 8.9
20:00 239.9 F 10.0
22:00 247.3 F 13.5
24:00 258.9 F 20.0
26:00 269.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.7 F 28.9
2:00 271.8 F 28.1
3:00 271.6 F 28.5
4:00 272.6 F 28.7

Min 271.0 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 210.1 F 0.0
10:00 227.7 F 0.0
15:00 222.2 F 0.0
20:00 225.8 F 0.0
25:00 235.2 F 0.0
30:00 225.9 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 27:02
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:05:09
CYCLE COMPLETE

10/07/2021 01:12 PM
ACCEPTED: X
REJECTED:
Operator: U02