
Midmark M11 - v1.0.5
Total Cycles: 4189

Name: RITTER2
Sterilizer ID: V2313228
12/05/2024 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.6 F 0.0
2:00 170.8 F 0.7
4:00 196.4 F 4.4
6:00 207.3 F 9.0
8:00 227.0 F 13.5
10:00 228.7 F 9.2
12:00 235.9 F 10.4
14:00 242.0 F 12.7
16:00 242.3 F 12.1
18:00 254.8 F 18.7
20:00 265.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.5 F 30.0
2:00 271.9 F 29.4
3:00 271.6 F 29.7
4:00 272.6 F 30.0

Min 271.2 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 230.7 F 0.0
10:00 225.9 F 0.0
15:00 226.5 F 0.0
20:00 230.8 F 0.0
25:00 230.8 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

12/05/2024 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____