
Midmark M11 - v1.0.5
Total Cycles: 3665

Name: RITTER2
Sterilizer ID: V2313228
06/04/2024 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 88.2 F 0.1
4:00 124.9 F 0.8
6:00 113.2 F 2.7
8:00 149.7 F 5.5
10:00 174.0 F 8.4
12:00 210.1 F 12.1
14:00 213.7 F 4.6
16:00 216.5 F 7.9
18:00 228.5 F 11.4
20:00 227.4 F 6.3
22:00 239.6 F 11.5
24:00 235.5 F 8.9
26:00 247.6 F 14.6
28:00 257.8 F 20.2
30:00 267.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.4 F 29.3
2:00 272.3 F 29.6
3:00 272.0 F 29.4
4:00 272.8 F 30.2

Min 271.0 F 28.8
Max 273.4 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 238.0 F 0.0
10:00 214.0 F 0.0
15:00 175.7 F 0.0
20:00 173.0 F 0.0
25:00 171.2 F 0.0
30:00 170.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

06/04/2024 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____