

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2547

Name: RITTER 1  
Sterilizer ID: V2178198  
11/11/2022 02:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.8 F 0.0  
2:00 88.5 F 0.0  
4:00 95.4 F 0.0  
6:00 116.5 F 1.5  
8:00 146.3 F 5.0  
10:00 183.2 F 8.8  
12:00 213.4 F 13.2  
14:00 215.8 F 6.1  
16:00 227.0 F 10.1  
18:00 234.4 F 8.9  
20:00 238.1 F 11.0  
22:00 234.6 F 8.4  
24:00 248.4 F 14.8  
26:00 259.8 F 21.1  
28:00 270.2 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.4 F 29.4  
2:00 271.5 F 28.5  
3:00 272.4 F 29.5  
4:00 271.8 F 28.8

Min 271.4 F 28.5  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 163.9 F 0.0  
10:00 198.5 F 0.0  
15:00 200.8 F 0.0  
20:00 205.3 F 0.0  
Dry Cycle Aborted

FILLING: 0:49  
HEATING: 28:55  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 20:21  
TOTAL CYCLE: 00:57:13

11/11/2022 03:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_