
Midmark M11 - v1.0.5
Total Cycles: 176

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.0 F 0.0
2:00 190.7 F 3.9
4:00 207.4 F 9.7
6:00 224.7 F 7.7
8:00 234.4 F 11.8
10:00 236.3 F 9.9
12:00 236.7 F 9.1
14:00 250.3 F 15.8
16:00 262.3 F 22.8
18:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.2 F 29.4
2:00 271.8 F 29.5
3:00 271.9 F 29.2
4:00 272.0 F 29.7

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 224.8 F 0.0
10:00 240.7 F 0.0
15:00 231.2 F 0.0
20:00 227.5 F 0.0
25:00 227.3 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 2:38
HEATING: 18:18
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

07/07/2021 03:17 PM
ACCEPTED: X
REJECTED:
Operator: U01