
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001458
11:46 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:16
STERILIZING 3:26
136.1 C 220kPa 3:26
136.4 C 226kPa 3:56
136.4 C 225kPa 4:26
136.4 C 223kPa 4:56
136.4 C 224kPa 5:26
136.3 C 225kPa 5:56
136.4 C 226kPa 6:26
136.4 C 225kPa 6:56
136.4 C 225kPa 6:56
VENTING 6:56
AIR DRYING 7:47
CYCLE COMPLETE 22:47

Digital Signature #
8269EE7437CE16AE
