
Midmark M11 - v1.0.5
Total Cycles: 4026

Name: RITTER2
Sterilizer ID: V2313228
10/11/2024 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.1 F 0.0
2:00 126.2 F 0.1
4:00 144.4 F 1.6
6:00 170.8 F 4.6
8:00 185.8 F 8.0
10:00 212.5 F 11.7
12:00 219.5 F 5.8
14:00 227.1 F 9.9
16:00 234.9 F 10.1
18:00 236.9 F 11.1
20:00 235.2 F 8.8
22:00 248.0 F 14.6
24:00 258.1 F 20.2
26:00 267.5 F 26.1
28:00 272.3 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 273.1 F 29.9
2:00 271.4 F 29.8
3:00 273.0 F 29.9
4:00 271.4 F 29.8

Min 271.2 F 29.5
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 227.8 F 0.0
10:00 190.0 F 0.0
15:00 187.7 F 0.0
20:00 185.1 F 0.0
25:00 183.6 F 0.0
30:00 181.3 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 28:29
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:07:22
CYCLE COMPLETE

10/11/2024 05:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____