

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3908

Name: RITTER 1  
Sterilizer ID: V2178198  
10/16/2023 01:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.0 F 0.0  
2:00 160.5 F 3.0  
4:00 174.0 F 7.0  
6:00 205.2 F 12.6  
8:00 218.7 F 6.6  
10:00 232.9 F 11.8  
12:00 232.8 F 8.5  
14:00 240.2 F 10.7  
16:00 245.5 F 13.3  
18:00 258.1 F 20.0  
20:00 269.8 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.2  
2:00 271.5 F 28.5  
3:00 272.7 F 29.4  
4:00 271.9 F 28.7

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 190.9 F 0.0  
10:00 224.3 F 0.0  
15:00 226.5 F 0.0  
20:00 229.3 F 0.0  
25:00 231.4 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 20:55  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

10/16/2023 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_