
Midmark M11 - v1.0.0
Total Cycles: 1729

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2022 11:25 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.0 F 0.0
2:00 134.5 F 0.9
4:00 160.6 F 4.4
6:00 189.3 F 8.8
8:00 207.3 F 6.2
10:00 224.7 F 8.2
12:00 236.6 F 10.2
14:00 241.1 F 11.6
16:00 239.2 F 9.8
18:00 253.3 F 16.9
20:00 265.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.0 F 28.2
2:00 272.7 F 28.7
3:00 272.7 F 29.0

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.8 F 0.0
Dry Cycle Aborted

FILLING: 1:31
HEATING: 21:37
STERILIZING: 3:00
VENTING: 2:49
DRYING: 0:26
TOTAL CYCLE: 00:29:23

03/21/2022 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____