

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3546

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2024 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 82.1 F 0.0

C103: HEATUP MODE

\*\* ITEMS NOT STERILE \*\*

04/22/2024 02:59 PM