

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002188  
8:11 01/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:06  
STERILIZING 3:20  
136.4 C 225kPa 3:20  
136.2 C 223kPa 3:50  
136.4 C 225kPa 4:20  
136.4 C 225kPa 4:50  
136.4 C 224kPa 5:20  
136.3 C 223kPa 5:50  
136.4 C 224kPa 6:20  
136.3 C 224kPa 6:50  
136.3 C 224kPa 6:50  
VENTING 6:51  
AIR DRYING 7:40  
CYCLE COMPLETE 37:40  
-----

Digital Signature #  
075E3DD75F5715E4  
-----