
Midmark M11 - v1.0.0
Total Cycles: 3767

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2023 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.7 F 0.0
2:00 143.9 F 0.0
4:00 165.0 F 1.7
6:00 189.1 F 4.4
8:00 219.6 F 8.0
10:00 232.7 F 11.9
12:00 225.3 F 6.1
14:00 237.5 F 10.5
16:00 234.3 F 8.2
18:00 244.5 F 12.7
20:00 242.1 F 11.2
22:00 254.1 F 17.2
24:00 264.6 F 23.5
26:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.7
2:00 271.5 F 28.2
3:00 272.5 F 28.7
4:00 271.6 F 28.1

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 225.7 F 0.0
10:00 220.1 F 0.0
15:00 207.8 F 0.0
20:00 202.7 F 0.0
25:00 199.8 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:01
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

09/13/2023 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____