
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000975
14:12 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:46
STERILIZING 3:55
136.4 C 233kPa 3:55
136.4 C 233kPa 4:25
136.4 C 233kPa 4:55
136.3 C 232kPa 5:25
136.2 C 231kPa 5:55
136.3 C 233kPa 6:25
136.3 C 231kPa 6:55
136.3 C 232kPa 7:25
136.3 C 232kPa 7:25
VENTING 7:25
AIR DRYING 8:17
CYCLE COMPLETE 13:42

Digital Signature #
DDA881B13713F2D4
