
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003643
8:57 11/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:12
STERILIZING 3:15
136.0 C 220kPa 3:15
136.2 C 223kPa 3:45
136.3 C 222kPa 4:15
136.4 C 223kPa 4:45
136.4 C 223kPa 5:15
136.1 C 222kPa 5:45
136.4 C 225kPa 6:15
136.2 C 221kPa 6:45
136.2 C 221kPa 6:45
VENTING 6:45
AIR DRYING 7:32
CYCLE COMPLETE 7:38

Digital Signature #
918EBDBC79F66D72
