

-----  
Midmark M11 - v1.0.5  
Total Cycles: 560

Name: RITTER2  
Sterilizer ID: V2313228  
11/17/2021 02:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.1 F 0.0  
2:00 171.0 F 3.4  
4:00 180.5 F 7.8  
6:00 203.9 F 13.0  
8:00 218.2 F 7.5  
10:00 232.3 F 12.8  
12:00 236.4 F 10.2  
14:00 235.3 F 8.7  
16:00 249.2 F 15.4  
18:00 260.8 F 22.1  
20:00 271.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.6 F 29.8  
2:00 271.7 F 29.1  
3:00 272.1 F 29.9  
4:00 272.2 F 29.6  
  
Min 271.3 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 229.4 F 0.0  
10:00 222.9 F 0.0  
15:00 230.3 F 0.0  
20:00 222.0 F 0.0  
25:00 223.5 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:39  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

11/17/2021 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_