
Midmark M11 - v1.0.5
Total Cycles: 1061

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.0 F 0.0
2:00 189.0 F 2.9
4:00 194.5 F 7.4
6:00 212.5 F 12.7
8:00 225.8 F 7.7
10:00 234.9 F 13.1
12:00 239.1 F 10.9
14:00 237.5 F 9.6
16:00 251.5 F 16.6
18:00 263.5 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 271.9 F 29.6
2:00 272.3 F 29.4
3:00 272.1 F 29.5
4:00 272.1 F 29.9

Min 270.9 F 28.8
Max 273.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 236.5 F 0.0
10:00 222.0 F 0.0
15:00 213.3 F 0.0
20:00 199.7 F 0.0
25:00 200.5 F 0.0
30:00 189.5 F 0.0

Drying Complete

FILLING: 1:44
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

05/03/2022 10:02 AM
ACCEPTED: X
REJECTED:
Operator: U02