

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2568

Name: RITTER 1  
Sterilizer ID: V2178198  
11/16/2022 01:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 175.5 F 4.3  
4:00 192.9 F 8.7  
6:00 211.9 F 10.8  
8:00 226.9 F 8.5  
10:00 240.3 F 11.7  
12:00 238.0 F 10.9  
14:00 236.7 F 8.9  
16:00 250.3 F 15.7  
18:00 262.1 F 22.4  
20:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.3 F 29.2  
2:00 271.4 F 28.6  
3:00 272.4 F 29.3  
4:00 271.8 F 28.7

Min 271.3 F 28.5  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 230.0 F 0.0  
10:00 236.9 F 0.0  
15:00 229.4 F 0.0  
20:00 231.7 F 0.0  
25:00 229.0 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:45  
CYCLE COMPLETE

11/16/2022 02:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_